

9.16.275E 27-03-2024

## Modbus RTU / FLOW-BUS



Pin 3 and 5 are internally connected. Pin 4 and 8 are internally connected.

i

For reliable measurements, it is important to reduce interference from outside noise on the signals.

- When connecting to other devices, ensure that the integrity of the shielding is maintained. ٠
- Use shielded cables and wire terminals.