


9.16.313A 27-03-2024


## Industrial Ethernet interfaces EtherCAT / EtherNet-IP / Modbus-TCP / POWERLINK / PROFINET


Hook-up diagram FLEXI-FLOW Compact Meter/Controller













Pin 3 and 5 are internally connected. Pin 4 and 8 are internally connected.



For reliable measurements, it is important to reduce interference from outside noise on the signals.

- When connecting to other devices, ensure that the integrity of the shielding is maintained.
- Use shielded cables and wire terminals.